

## Determination of the Parameters of Prototype of “Gamma” Installation at the IREN Facility

*Tuesday 2 July 2024 13:15 (15 minutes)*

On the flight bases of the 4th (11 meters) channels of the IREN facility, neutron flux density measurements were performed in the neutron energy range from 0.4 eV to 1500 eV. The neutron fluxes (thermal and resonance) were determined using gold foils. The parameters of the resolution function have also been determined and their values, obtained from experimental measurements for several neutron energies, are in good agreement with the calculations.

### Section

Applications of nuclear methods in science, technology, medicine and radioecology

**Primary author:** YERGASHOV, Almat (Frank Laboratory of Neutron Physics, Joint Institute for Nuclear Research, Institute of Nuclear Physics, Almaty, Republic of Kazakhstan)

**Co-authors:** Mr HRAMCO, Constantin (Frank Laboratory of Neutron Physics, Joint Institute for Nuclear Research); Dr MITSYNA, Ludmila; Dr REBROVA, Natalia (Frank Laboratory of Neutron Physics, Joint Institute for Nuclear Research); Mr SEDYSHEV, Pavel (Frank Laboratory of Neutron Physics, Joint Institute for Nuclear Research); BORZAKOV, Sergey (FLNP JINR); Dr KUZNETSOV, Valery (Frank Laboratory of Neutron Physics, Joint Institute for Nuclear Research); KOPATCH, Yuri (Joint Institute for Nuclear Research)

**Presenter:** YERGASHOV, Almat (Frank Laboratory of Neutron Physics, Joint Institute for Nuclear Research, Institute of Nuclear Physics, Almaty, Republic of Kazakhstan)

**Session Classification:** Applications of nuclear methods in science, technology, medicine and radioecology