

NEC'2019



Contribution ID: 125

Type: **Sectional**

Electronics of the fission fragments spectrometer "COMETA-F"

Thursday 3 October 2019 15:15 (15 minutes)

The article describes the electronics for the time-of-flight two-arm fission fragments spectrometer COMETA-F. The time pick-off detector is comprised of thin electron conversion foil, electrostatic mirror and of two microchannel plates, supplied by Baspik, mounted in a chevron configuration. Mosaics of Si PiN diodes used to measure both energy and time-of-flight. The waveform of detected signals is digitized by V1742 modules, which sample signals through the DRS4 chip. The DRS4 chip is a switched capacitor array, which can sample the input signal at a frequency of 5 GHz. Start for registration is provided by a specially designed trigger module. The V945 discriminator thresholds are individually settable in a range from -1 mV to -255 mV via VME through an 8-bit DAC.

The use of the Si-semiconductor detectors in time-of-flight-energy spectrometry of fission fragments is known to have delicate methodological problems due to the "amplitude (pulse-height) defect" and "plasma delay" effects in the E and TOF channels, respectively. Correct accounting for both effects needs rather complicated procedure of the mass reconstruction. Using off-line algorithm and mass reconstruction procedure based on the PHD parametrization let us to reproduce quite satisfactory ion masses in a wide range of masses and energies.

Primary author: Dr STREKALOVSKY, Oleg (JINR FLNR)

Co-authors: Mr STREKALOVSKY, Alexandr (JINR); KAMANIN, Dmitry (Joint Institute for Nuclear Research); Prof. PYATKOV, Yury (National Nuclear Research University "MEPHI" & Joint Institute for Nuclear Research)

Presenter: Dr STREKALOVSKY, Oleg (JINR FLNR)

Session Classification: Triggering, Data Acquisition, Control Systems

Track Classification: Triggering, Data Acquisition, Control Systems